AIV

AGGREGATE

BS 812

- Used to determine the aggregate particle resistance to impact.
- Supplied with automatic blow counter.
- Supplied with a sample cylinder, measuring cylinder and a tamping rod.



SUPPLIED WITH

- Cylindrical measure (Ø 75 mm x 50 mm)
- Tamping Rod

EQUIPPED WITH

- Blow counter
- Hammer release mechanism

established in 1972 www.alfatestusa.com